



PCT

WORLD INTELLECTUAL PROPERTY ORGANIZATION  
International Bureau

## INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

<p>(51) International Patent Classification 7 : <b>G01N 27/64, H01J 49/04, 49/42, B01D 59/48</b></p>	<p><b>A1</b></p>	<p>(11) International Publication Number: <b>WO 00/08454</b> (43) International Publication Date: 17 February 2000 (17.02.00)</p>
<p>(21) International Application Number: PCT/CA99/00714 (22) International Filing Date: 5 August 1999 (05.08.99) (30) Priority Data: 60/095,481 5 August 1998 (05.08.98) US 2,260,572 29 January 1999 (29.01.99) CA 09/321,820 28 May 1999 (28.05.99) US (71) Applicant (for all designated States except US): NATIONAL RESEARCH COUNCIL CANADA [CA/CA]; 1500 Montreal Road, Ottawa, Ontario K1A 0R6 (CA). (72) Inventors; and (75) Inventors/Applicants (for US only): GUEVREMONT, Roger [CA/CA]; 2059 Gatineau View Cr., Gloucester, Ontario K1J 7W9 (CA). PURVES, Randy, W. [CA/CA]; 59-6247 Sundown Cr., Gloucester, Ontario K1C 2M1 (CA). BARNETT, David [CA/CA]; 1934 Longman Cr., Orleans, Ontario K1C 5G6 (CA). (74) Agent: BERESKIN &amp; PARR; 40th floor, 40 King Street West, Toronto, Ontario M5H 3Y2 (CA).</p>		<p>(81) Designated States: AE, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CR, CU, CZ, DE, DK, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, UA, UG, US, UZ, VN, YU, ZA, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).</p> <p>Published With international search report. Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.</p>
<p>(54) Title: METHOD FOR SEPARATION OF ISOMERS AND DIFFERENT CONFORMATIONS OF IONS IN GASEOUS PHASE</p> <p>(57) Abstract</p> <p>The present invention relates to a method of separating ions having the same mass to charge ratio but different ion mobility characteristics, based on the ion focussing principles of high field asymmetric waveform ion mobility spectrometry. Isomers including gas-phase molecular anions of the amino acids leucine and isoleucine are separable by the method of the present invention. Identification of different conformers of ions at the same charge state, including conformers present at certain charge states of the protein bovine ubiquitin, is also possible using the method of the present invention.</p>		

AR

**FOR THE PURPOSES OF INFORMATION ONLY**

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece	ML	Mali	TR	Turkey
BG	Bulgaria	HU	Hungary	MN	Mongolia	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MR	Mauritania	UA	Ukraine
BR	Brazil	IL	Israel	MW	Malawi	UG	Uganda
BY	Belarus	IS	Iceland	MX	Mexico	US	United States of America
CA	Canada	IT	Italy	NE	Niger	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NL	Netherlands	VN	Viet Nam
CG	Congo	KE	Kenya	NO	Norway	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NZ	New Zealand	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	PL	Poland		
CM	Cameroon	KR	Republic of Korea	PT	Portugal		
CN	China	KZ	Kazakhstan	RO	Romania		
CU	Cuba	LC	Saint Lucia	RU	Russian Federation		
CZ	Czech Republic	LI	Liechtenstein	SD	Sudan		
DE	Germany	LK	Sri Lanka	SE	Sweden		
DK	Denmark	LR	Liberia	SG	Singapore		
EE	Estonia						

## METHOD FOR SEPARATION OF ISOMERS AND DIFFERENT CONFORMATIONS OF IONS IN GASEOUS PHASE

FIELD OF THE INVENTION

The present invention relates to a method for separating isomers and  
5 different conformations of ions in gaseous phase, based on the principle of high field  
asymmetric waveform ion mobility spectrometry.

BACKGROUND OF THE INVENTION

High sensitivity and amenability to miniaturization for field-portable  
applications have helped to make ion mobility spectrometry an important technique  
10 for the detection of many compounds, including narcotics, explosives, and chemical  
warfare agents (see, for example, G. Eiceman and Z. Karpas, *Ion Mobility  
Spectrometry* (CRC, Boca Raton, FL, 1994); and *Plasma Chromatography*, edited by  
T.W. Carr (Plenum, New York, 1984)). In ion mobility spectrometry, gas-phase ion  
mobilities are determined using a drift tube with a constant electric field. Ions are  
15 gated into the drift tube and are subsequently separated based upon differences in  
their drift velocity. The ion drift velocity is proportional to the electric field strength  
at low electric fields (e.g., 200 V/cm) and the mobility,  $K$ , which is determined from  
experimentation, is independent of the applied field. At high electric fields (e.g.  
5000 or 10000 V/cm), the ion drift velocity may no longer be directly proportional to  
20 the applied field, and  $K$  becomes dependent upon the applied electric field (see G.  
Eiceman and Z. Karpas, *Ion Mobility Spectrometry* (CRC, Boca Raton, FL, 1994); and  
E.A. Mason and E.W. McDaniel, *Transport Properties of Ions in Gases* (Wiley, New  
York, 1988)). At high electric fields,  $K$  is better represented by  $K_h$ , a non-constant  
high field mobility term. The dependence of  $K_h$  on the applied electric field has been  
25 the basis for the development of high field asymmetric waveform ion mobility  
spectrometry (FAIMS), a term used by the inventors throughout this disclosure, and  
also referred to as transverse field compensation ion mobility spectrometry, or field  
ion spectrometry (see I. Buryakov, E. Krylov, E. Nazarov, and U. Rasulev, *Int. J.  
Mass Spectrom. Ion Proc.* 128, 143 (1993); D. Riegner, C. Harden, B. Carnahan, and  
30 S. Day, *Proceedings of the 45th ASMS Conference on Mass Spectrometry and Allied  
Topics*, Palm Springs, California, 1-5 June 1997, p. 473; B. Carnahan, S. Day, V.

Kouznetsov, M. Matyjaszczyk, and A. Tarassov, Proceedings of the 41st ISA Analysis Division Symposium, Framingham, MA, 21-24 April 1996, p. 85; and B. Carnahan and A. Tarassov, U.S. Patent Number 5,420,424). Ions are separated in FAIMS on the basis of the difference in the mobility of an ion at high field  $K_h$  relative to its mobility at low field  $K$ . That is, the ions are separated because of the compound dependent behaviour of  $K_h$  as a function of the electric field. This offers a new tool for atmospheric pressure gas-phase ion studies since it is the change in ion mobility and not the absolute ion mobility that is being monitored.

An instrument based on the FAIMS concept has been designed and built by Mine Safety Appliances Company of Pittsburgh, Pa. ("MSA") for use in trace gas analysis. The MSA instrument is described in U.S. Patent No. 5,420,424 and is available under the trade mark FIS (for Field Ion Spectrometer). While the use of the MSA instrument (and similar instruments based on the FAIMS concept) for trace gas analysis is known, the inventors believe that they have identified certain heretofore unrealized properties of these instruments which make them more versatile. Based on this realization, the inventors have developed what is believed to be a previously unknown method for separation of isomers and different conformations of ions. A summary and detailed description of the present invention is provided below.

## 20 SUMMARY OF THE INVENTION

The present invention provides a method for identifying ions having substantially the same mass to charge ratio but having different ion mobility characteristics, comprising the steps of:

- a) providing at least one ionization source of ions;
- 25 b) providing an analyzer region defined by a space between at least first and second spaced apart electrodes, said analyzer region being in communication with a gas inlet, a gas outlet, an ion inlet and an ion outlet, and introducing said ions into said analyzer region through said ion inlet;
- 30 c) applying an asymmetric waveform voltage and a direct current

- 3 -

- compensation voltage to at least one of said electrodes;
- d) setting said asymmetric waveform voltage;
  - e) varying said direct current compensation voltage and measuring resulting transmitted ions at said ion outlet, so as to produce a compensation voltage scan for said transmitted ions; and
  - f) identifying peaks in said compensation voltage scan.

The method may further comprise the step of setting said direct current compensation voltage to correspond to one of said peaks to separate a desired ion from other ions with substantially the same mass to charge ratio.

Advantageously, the above method is operable at substantially at atmospheric pressure and substantially at room temperature.

The method may further include detecting said transmitted ions by mass spectrometry.

Typically, the method includes providing a gas flow through said analyzer region, so as to transport said ions along said analyzer region, although it will be understood that other ion transport means are possible.

Furthermore, in identifying a peak, it will be understood that the term peak is not limited to the apex of the peak, and that a peak will typically have a noticeable width, or a compensation voltage range in which the peak appears.

## **BRIEF DESCRIPTION OF THE DRAWINGS**

For a better understanding of the present invention, and by way of example, reference will now be made to the accompanying drawings, which show preferred embodiments of the present invention in which:

Figure 1 shows three possible examples of changes in ion mobility as a function of the strength of an electric field;

Figure 2 illustrates the trajectory of an ion between two parallel plate electrodes under the influence of the electrical potential  $V(t)$ ;

Figures 3A and 3B show schematically an embodiment of a modified FAIMS device;

Figure 4 illustrates two opposite waveform modes which may be used

with the apparatus of Figures 3A and 3B;

Figures 5A and 5B show schematically the coupling of the FAIMS apparatus of Figures 3A and 3B together with a mass spectrometer;

Figures 6A and 6B shows schematically a FAIMS apparatus for  
5 measuring the ion distribution in the analyzer region;

Figures 7 illustrates the high voltage, high frequency asymmetric waveform applied to the FAIMS apparatus shown in Figures 6A and 6B;

Figure 8 illustrates varying ion arrival time profiles at the innermost ion collector electrode of the FAIMS apparatus in Figures 6A and 6B;

10 Figure 9 shows an ion selected compensation voltage (IS-CV) spectra at five different DV values for a leucine/isoleucine mixture;

Figure 10 trace (a) shows the IS-CV spectra, run separately, of a solution containing leucine and a solution containing isoleucine;

Figure 10 trace (b) shows an IS-CV spectrum of a leucine/isoleucine  
15 mixture;

Figure 11A shows mass spectra for a leucine/isoleucine mixture before filtering through a FAIMS analyzer;

Figures 11B and 11C show mass spectra for a leucine/isoleucine mixture after filtering through a FAIMS analyzer at two different CV values;

20 Figure 12 shows a response curve for leucine plotted as a function of concentration;

Figure 13 shows an expanded view of an IS-CV spectrum acquired for a solution containing 0.004  $\mu$ M leucine and 2.496  $\mu$ M isoleucine;

Figure 14A shows an ESI mass spectrum for a solution of bovine  
25 ubiquitin;

Figure 14B shows a total ion current CV (TIC-CV) spectrum of a solution of bovine ubiquitin;

Figures 14C-14E show mass spectra obtained at several different CV values;

30 Figure 15 shows normalized IS-CV spectra for various charge states of bovine ubiquitin ranging from +5 to +13 using the same solution as that used for Figure 14;

Figures 16A, 16C, and 16E show mass spectra showing the effect of the amount of acetic acid on ESI mass spectra of a solution of bovine ubiquitin;

Figures 16B, 16D, and 16F show TIC-CV spectra corresponding to the ESI-mass spectra obtained in Figures 16A, 16C, and 16E, respectively;

5        Figures 17A-17I show IS-CV spectra showing the effect of the amount of acetic acid in a solution of bovine ubiquitin on the charge states +7, +8 and +9;

Figures 18A-18I show IS-CV spectra showing the effect of the amount of HCl in a solution of bovine ubiquitin on the charge states +7, +8 and +9;

Figure 19 shows normalized IS-CV spectra for the charge states +5 to  
10        +13 using a 5  $\mu$ M solution of bovine ubiquitin (55% water) acidified to pH 2.1 using HCl;

Figures 20A-20F show the effect of solvent composition on mass spectra, TIC-CV spectra, and IS-CV spectra of bovine ubiquitin;

Figures 21A and 21B show the effect of adding NaCl to a solution of  
15        bovine ubiquitin on the IS-CV spectrum for charge state +8;

Figures 21C through 21H show mass spectra of different CV values;

Figure 22A shows an IS-CV spectrum showing the dependence of sodium adduct ion intensity on the conformation for the +6 charge state of bovine ubiquitin;

20        Figures 22B and 22C show mass spectra for the solution used in Figure 21A at two different CV values;

Figure 23A shows an IS-CV spectrum for the +8 charge state of bovine ubiquitin using a solution containing phosphate;

Figures 23B-23D show mass spectra for the solution used in Figure 23A  
25        at three different CV values; and

Figure 24 is a plot showing the location of the peak maxima for all conformers of bovine ubiquitin observed in this study.

#### DETAILED DESCRIPTION OF THE INVENTION

30        As an important preliminary note, the discussion below generally uses the term "ion" to mean a charged atomic or molecular entity, the "ion" can be any electrically charged particle, solid or liquid, of any size. The discussion below refers

to both positively charged and negatively charged ions, and it will be understood by a person skilled in the art that, for any individual analysis, only one of these types of ions will be used.

The discussion below also uses the term "isomers" to mean compounds  
5 having identical molecular formulas but which differ in the ways in which the atoms are bonded to each other. Generally speaking, isomers may be constitutional isomers or stereoisomers. Constitutional isomers differ in the order and the way in which atoms are bonded together in their molecules. Stereoisomers differ only in the arrangement of their atoms in space. Stereoisomers that are nonsuperimposable  
10 mirror images of each other are called enantiomers. Stereoisomers that are not enantiomers are called diastereomers. Also, the lack of free rotation around carbon-carbon bonds may form cis-trans isomers where two substituents may be on opposite sides of a plane (trans) or on the same side of a plane (cis). Finally, positional isomers (e.g. ortho, meta, and para positions within a carbon ring) and  
15 geometrical isomers may form in various types of compounds.

The disclosure also uses the term "ion selected compensation voltage" (IS-CV) spectra which refers to scanning the compensation voltage applied to a FAIMS analyzer, as discussed below, typically while monitoring a single mass-to-charge ( $m/z$ ) value. The term "total ion current compensation voltage" (TIC-CV)  
20 spectra is also used to refer to a compensation voltage scan which shows the sum of a signal for all detected ions in a given  $m/z$  range.

#### *Principles of FAIMS*

The principles of operation of FAIMS have been described in Buryakov *et. al.* (see I. Buryakov, E. Krylov, E. Nazarov, and U. Rasulev, *Int. J. Mass Spectrom. Ion Proc.* 128. 143 (1993)) and are summarized here briefly. The mobility of a given ion under the influence of an electric field can be expressed by:  $K_h(E) = K(1+f(E))$ , where  $K_h$  is the mobility of an ion at high field,  $K$  is the coefficient of ion mobility at low electric field and " $f(E)$ " describes the functional dependence of the ion mobility on the electric field (see E.A. Mason and E.W. McDaniel, *Transport Properties of Ions in Gases* (Wiley, New York, 1988); and I. Buryakov, E. Krylov, E. Nazarov, and U. Rasulev, *Int. J. Mass Spectrom. Ion Proc.* 128. 143 (1993)).  
25  
30

Referring to Figure 1, three examples of changes in ion mobility as a function of the strength of an electric field are shown: the mobility of type A ions increases with increasing electric field strength; the mobility of type C ions decreases; and the mobility of type B ions increases initially before decreasing at yet higher fields. The separation of ions in FAIMS is based upon these changes in mobility at high electric fields. Consider an ion 1, for example a type A ion shown in Figure 1, that is being carried by a gas stream 6 between two spaced apart parallel plate electrodes 2, 4 as shown in Figure 2. The space between the plates 2, 4 defines an analyzer region 5 in which the separation of ions may take place. The net motion of the ion 1 between the plates 2, 4 is the sum of a horizontal x-axis component due to a flowing stream of gas 6 and a transverse y-axis component due to the electric field between the plates 2, 4. (The term "net" motion refers to the overall translation that the ion 1 experiences, even when this translational motion has a more rapid oscillation superimposed upon it.) One of the plates is maintained at ground potential (here, the lower plate 4) while the other (here, the upper plate 2) has an asymmetric waveform,  $V(t)$ , applied to it. The asymmetric waveform  $V(t)$  is composed of a high voltage component,  $V_1$ , lasting for a short period of time  $t_2$  and a lower voltage component,  $V_2$ , of opposite polarity, lasting a longer period of time  $t_1$ . The waveform is synthesized such that the integrated voltage-time product (thus the field-time product) applied to the plate during a complete cycle of the waveform is zero (i.e.,  $V_1 t_2 + V_2 t_1 = 0$ ); for example +2000 V for 10  $\mu$ s followed by -1000 V for 20  $\mu$ s. Figure 2 illustrates the ion trajectory 8 (as a dashed line) for a portion of the waveform shown as  $V(t)$ . The peak voltage during the shorter, high voltage portion of the waveform will be called the "dispersion voltage" or DV in this disclosure.

During the high voltage portion of the waveform, the electric field will cause the ion 1 to move with a transverse velocity component  $v_1 = K_h E_{\text{high}}$ , where  $E_{\text{high}}$  is the applied field, and  $K_h$  is the high field mobility under ambient electric field, pressure and temperature conditions. The distance travelled will be  $d_1 = v_1 t_2 = K_h E_{\text{high}} t_2$ , where  $t_2$  is the time period of the applied high voltage. During the longer duration, opposite polarity, low voltage portion of the waveform, the velocity component of the ion will be  $v_2 = K E_{\text{low}}$ , where  $K$  is the low field ion mobility under ambient pressure and temperature conditions. The distance travelled is  $d_2 = v_2 t_1 = K E_{\text{low}} t_1$ .

Since the asymmetric waveform ensures that  $(V_1 t_2) + (V_2 t_1) = 0$ , the field-time products  $E_{\text{high}} t_2$  and  $E_{\text{low}} t_1$  are equal in magnitude. Thus, if  $K_h$  and  $K$  are identical,  $d_1$  and  $d_2$  are equal, and the ion 1 will be returned to its original position along the y-axis during the negative cycle of the waveform (as would be expected if both portions of the waveform were low voltage). If at  $E_{\text{high}}$  the mobility  $K_h > K$ , the ion 1 will experience a net displacement from its original position relative to the y-axis. For example, positive ions of the type A shown in Figure 1 will travel further during the positive portion of the waveform (i.e.,  $d_1 > d_2$ ) and the type A ion 1 will migrate away from the upper plate 2 (as illustrated by the dashed line 8 in Figure 2). Similarly, ions of type C will migrate towards the upper plate 2.

If an ion of type A is migrating away from the upper plate 2, a constant negative dc voltage can be applied to this plate 2 to reverse, or "compensate" for this transverse drift. This dc voltage, called the "compensation voltage" or CV in this disclosure, prevents the ion 1 from migrating towards either plate 2, 4. If ions derived from two compounds respond differently to the applied high electric fields, the ratio of  $K_h$  to  $K$  may be different for each compound. Consequently, the magnitude of the compensation voltage CV necessary to prevent the drift of the ion toward either plate 2, 4 may also be different for each compound. Under conditions in which the compensation voltage CV is appropriate for transmission of one compound, the other will drift towards one of the plates 2, 4 and subsequently be lost. The speed at which the compound will move to the wall of the plates 2, 4 depends on the degree to which its high field mobility properties differ from those of the compound that will be allowed to pass under the selected condition. A FAIMS instrument or apparatus is an ion filter capable of selective transmission of only those ions with the appropriate ratio of  $K_h$  to  $K$ .

The term FAIMS, as used in this disclosure, refers to any device which can separate ions via the above described mechanism, whether or not the device has focussing or trapping behaviour.

#### *Improvements to FAIMS*

The FAIMS concept was first shown by Buryakov *et. al.* using flat plates as described above. Later, Carnahan *et. al.* improved the sensor design by replacing

the flat plates used to separate the ions with concentric cylinders (see B. Carnahan, S. Day, V. Kouznetsov, M. Matyjaszczyk, and A. Tarassov, Proceedings of the 41st ISA Analysis Division Symposium, Framingham, MA, 21-24 April 1996, p. 85; U.S. Patent No. 5,420,424 issued to Carnahan et al.). The concentric cylinder design has several advantages including higher sensitivity than the flat plate configuration (see  
5 R.W. Purves, R. Guevremont, S. Day, C.W. Pipich, and M.S. Matyjaszczyk, Rev. Sci. Instrum., 69, 4094 (1998)).

As mentioned earlier, an instrument based on the FAIMS concept has been built by Mine Safety Appliances Company (MSA). The MSA instrument uses  
10 the concentric cylinder design and is described further below. (For the purposes of this disclosure, the MSA instrument is referred to as FAIMS-E, where E refers to an electrometer or electric current detection device.)

One previous limitation of the cylindrical FAIMS technology (see D. Riegner, C. Harden, B. Carnahan, and S. Day, Proceedings of the 45th ASMS  
15 Conference on Mass Spectrometry and Allied Topics, Palm Springs, California, 1-5 June 1997, p. 473; and B. Carnahan, S. Day, V. Kouznetsov, M. Matyjaszczyk, and A. Tarassov, Proceedings of the 41st ISA Analysis Division Symposium, Framingham, MA, 21-24 April 1996, p. 85) was that the identity of the peaks appearing in the FAIMS-E CV spectra could not be unambiguously confirmed due to  
20 the unpredictable changes in  $K_n$  at high electric fields.

Thus, one way to extend the capability of instruments based on the FAIMS concept, such as the FAIMS-E instrument, is to provide a way to determine the make-up of the FAIMS-E CV spectra more accurately, for example, by introducing ions from the FAIMS-E device into a mass spectrometer for mass-to-  
25 charge ( $m/z$ ) analysis.

In addition, it has been found that a modified FAIMS instrument, or any similar instrument, can be used in a new method of separating isomers and different conformations of gaseous phase ions. The present invention is directed to a new method of separating isomers and different conformations of ions and illustrates the  
30 method by several examples. Details of the method of the present invention are described below.

### *Electrospray Ionization*

ESI is one of several related techniques that involves the transfer of ions (which can be either positively or negatively charged) from liquid phase into the gas-phase. Kebarle has described four major processes that occur in electrospray ionization (intended for use in mass spectrometry): (1) production of charged droplets, (2) shrinkage of charged droplets by evaporation, (3) droplet disintegration (fission), and (4) formation of gas-phase ions (Kebarle, P. and Tang, L. *Analytical Chemistry*, 65 (1993) pp. 972A-986A). In ESI, a liquid solution (e.g. 50/50 w/w water/methanol) is passed through a metal capillary (e.g., 200  $\mu\text{m}$  outer diameter and 100  $\mu\text{m}$  ID) which is maintained at a high voltage to generate the charged droplets, say +2000 V (50 nA) for example. The liquid samples can be pumped through at, say, 1  $\mu\text{L}/\text{min}$ . The high voltage creates a very strong, non-constant electric field at the exit end of the capillary, which nebulizes the liquid exiting from the capillary into small charged droplets and electrically charged ions by mechanisms described by Kebarle and many others. Several related methods also exist for creating gas-phase ions from solution phase. Some examples of these methods include ionspray, which uses mechanical energy from a high velocity gas to assist in nebulization; thermospray, which applies heat instead of a voltage to the capillary; and nanospray, which uses small ID capillaries. In this disclosure, the term ESI is used to encompass any technique that creates gas-phase ions from solution.

### *Modified FAIMS-E*

As a first step, the FAIMS-E device designed and built by Mine Safety Appliances Company was modified to permit the introduction of ions using ESI. The inventors believe that the coupling of an ESI source together with a FAIMS-E device is not obvious as it is known that ions produced by ESI have a high degree of solvation, and that a FAIMS-E device may not function properly when exposed to high levels of solvent vapour. The inventors have developed various practical embodiments of an apparatus that combines an ESI source together with a FAIMS device to show that such coupling is possible.

One example is the modified FAIMS-E device 10 shown schematically in 3-dimensional view in Figure 3A and in cross section in Figure 3B. The FAIMS-E

- 11 -

apparatus 10 is composed of two short inner cylinders or tubes 11, 12 which are axially aligned and positioned about 5 mm apart, and a long outer cylinder 13 which surrounds the two inner cylinders 11, 12. The inner cylinders 11, 12 (12 mm inner diameter, 14 mm outer diameter), are about 30 mm and 90 mm long, respectively, while the outer cylinder 13 (18 mm inner diameter, 20 mm outer diameter) is about 125 mm long. Ion separation takes place in the 2 mm annular space of FAIMS analyzer region 14 between the long inner cylinder 12 and the outer cylinder 13. To produce ions using electrospray ionization (ESI), for introduction into the FAIMS analyzer region 14 of the FAIMS device, the metal capillary of the ESI needle 15 was placed along the central axis of the shorter inner cylinder 11, terminating about 5 mm short of the gap or ion inlet<sup>11A</sup> between the two inner cylinders 11, 12. The positioning of the ESI needle 15 shown in Figures 3(A) and 3(B) differs from the positioning of the ionization source found in the MSA FAIMS-E device in that the ESI needle 15 does not extend through the long inner cylinder 12 to which the asymmetric waveform V(t) is typically applied. By introducing the ESI needle 15 from the opposite end of the FAIMS-E, i.e. through the short inner cylinder 11, and not positioning the tip of the ESI needle 15 too close to the long inner cylinder 12, the performance of the ESI needle 15 is not compromised by the asymmetric waveform V(t), which would be the case if the ESI needle 15 was positioned within the long inner cylinder 12 (as disclosed in U.S. Patent No. 5,420,424).

As explained above, the FAIMS-E device 10 can be considered as an ion "filter", with the capability of selectively transmitting one type of ion out of a mixture. If a mixture of ions is presented continuously to the entrance of the FAIMS analyzer region 14, for example by an ESI needle 15, and the ions are carried along the length of the analyzer 14 by a flowing gas under conditions in which no voltages are applied to either the inner cylinder 12 or outer cylinder 13 (i.e. the electrodes are grounded), some finite level of transmission for every ion is expected, albeit without any separation.

It might be expected that the detected current of any selected ion in this mixture should never exceed the current for that ion when it is transmitted through the device 10 in the no-voltages condition. It might also be expected that application of high voltages (i.e. application of transverse fields, perpendicular to the gas flows)

designed to yield ion separation should not increase the ion transmission, but should decrease transmission through collisions with the walls of the cylinders 12, 13. That is, the asymmetric waveform might effectively narrow the "width" of the FAIMS analyzer region 14, and therefore should decrease the ion transmission. However, 5 contrary to this prediction, experiments conducted by the inventors and described in this disclosure have shown that the sensitivity of ion detection in the cylindrical geometry FAIMS-E 10 increases as the voltage amplitude of the asymmetric waveform  $V(t)$  is increased. As will be explained below, these unusual observations suggest that atmospheric pressure ion focussing is occurring in the FAIMS analyzer 10 region 14.

Still referring to Figures 3A and 3B, four gas connections to the FAIMS-E apparatus 10 are shown. Compressed gas (e.g. air or nitrogen) is passed through a charcoal/molecular sieve gas purification cylinder (not shown) into the FAIMS-E 10 through carrier in ( $C_{in}$ ) and/or sample in ( $S_{in}$ ) ports. The gas exits the FAIMS-E 10 15 via the carrier out ( $C_{out}$ ) and/or sample out ( $S_{out}$ ) ports. All four gas flow rates can be adjusted. Non-volatile analytes are typically introduced into the FAIMS-E 10 using an ESI needle 15. Alternatively, volatile analytes may be introduced into the FAIMS-E 10 through the  $S_{in}$  line, and a portion may be ionized as the compound(s) pass by a corona discharge needle.

20 Still referring to Figures 3A and 3B, the outer cylinder 13 of the FAIMS-E apparatus 10, and the shorter inner cylinder 11, are typically held at an adjustable electrical potential ( $V_{FAIMS}$ ).  $V_{FAIMS}$  is usually ground potential in FAIMS-E. During operation, a high frequency high voltage asymmetric waveform is applied to the long inner cylinder 12 to establish the electric fields between the inner and outer cylinders 25 12, 13. In addition to this high frequency (e.g., 210 kHz) high voltage waveform a dc offset voltage (i.e. the compensation voltage CV added to FAIMS) is applied to the long inner cylinder 12. This leads to the separation of ions in the FAIMS analyzer region 14 in the manner discussed earlier.

Still referring to Figures 3A and 3B, some of the ions produced by the 30 ionization source are carried by the gas stream along the length of the annular space between the outer cylinder 13 and the long inner cylinder 12, also referred to as the FAIMS analyzer region 14. If the combination of DV and CV are appropriate, and

PC

RF

- 13 -

the ion is not lost to the tube walls, a series of openings or ion outlets 16 near the downstream end of the outer cylinder 13 allow the ions to be extracted to an electrical current detector 17 which is biased to about -100 V. (Note that here the carrier gas also exits from the ion outlet 16.)

5           In practice, the simplified square wave version of  $V(t)$  shown in Figure 2 cannot be used because of the electrical power demands that such a wave would place on the waveform generator. The actual waveforms  $V(t)$  appear in Figure 4. These waveforms are produced by the electronic addition of a sine wave and its harmonic of twice the frequency. As shown in Figure 4, the FAIMS-E apparatus 10 operates using one of the two waveform modes (with the waveform applied to the inner cylinder). These reversed polarity waveform modes do not yield "reversed polarity" CV spectra as might be expected. This is because the reversal of polarity in this manner also creates a mirror image effect of the ion focussing behaviour of FAIMS. The result of such polarity reversal is that the ions are not focussed, but 15 rather collide with the walls of the cylinders 12, 13. The mirror image of a focussing valley is a hill-shaped potential surface. (This characteristic, and the various "modes" of operation of FAIMS, is discussed further below.)

#### *FAIMS-MS*

As discussed earlier, one way to extend the functionality of FAIMS 20 devices is to couple them together with a mass spectrometer. The use of a mass spectrometer together with a FAIMS device is advantageous because the mass spectrometer facilitates a mass-to-charge ( $m/z$ ) analysis to determine the make-up of CV spectra more accurately. One possible FAIMS-MS embodiment is described here.

25           Referring to Figures 5A and 5B, the coupling of FAIMS and a mass spectrometer (FAIMS-MS 20) is shown schematically. The FAIMS-MS 20 of Figures 5A and 5B, and the FAIMS-E 10 shown in Figures 3A and 3B, differ significantly only at the detection end of the instrument. In accordance with the invention, the electrometer 17 has been replaced by a sampler cone 18, placed at the end of the 30 FAIMS cylinders 12, 13 as is shown in a simplified form in Figure 5B. The diameter of the orifice 19 in the sampler cone 18 is approximately 250  $\mu\text{m}$ . The gas flows in

the FAIMS-MS 20 are analogous to those in the FAIMS-E 10 except that the  $C_{out}$  is divided into two components, namely the original  $C_{out}$  and the flow through the orifice 19 into the mass spectrometer. The electrical waveforms applied to the long inner cylinder 12 are identical to those used in the FAIMS-E apparatus 10. The  
5 sampler cone 18 may be electrically insulated from the other components so a separate voltage OR can be applied to it. Furthermore, a voltage can be applied to the cylinders of the entire FAIMS unit ( $V_{FAIMS}$ ) for the purpose of enhancing the sensitivity of the FAIMS-MS.

Figure 5B shows the FAIMS cylinders 12, 13 at a 45 degree angle in  
10 relation to the sampler cone 18 of the mass spectrometer. Figure 5A showed the FAIMS cylinders 12, 13 at a 90 degree angle in relation to the sampler cone 18. The way (i.e., the angle between the two tubes of the FAIMS and the sampler cone 18) in which the ions are extracted from the cylinders 12, 13 of the FAIMS-MS 20 into the mass spectrometer is not limited to these angles. Furthermore, the location in which  
15 the ions are extracted from the two tubes can also be changed. That is, the ions can be extracted anywhere along the separation region of the FAIMS.

#### ***Ion Focussing***

Referring now to Figures 6A and 6B, to demonstrate the focussing effect referred to above, a special FAIMS instrument was designed by the inventors and  
20 constructed to measure the ion distribution between the two cylinders (outer and inner cylinders) of a FAIMS device. This instrument will be referred to in this disclosure as the FAIMS-R1-prototype 30 and is illustrated schematically in Figures 6A and 6B. Ions were generated inside of an electrically grounded cylinder 31 approximately 35 mm long and 20 mm i.d.. The tip of an ionization needle 15 was  
25 typically located near the center of this tube, and at least 15 mm from the end of the FAIMS analyzer region 34. The FAIMS analyzer region 34 in this embodiment is composed of an outer tube 32 which is 70 mm long and 6 mm i.d., and which surrounds a 2 mm o.d. inner shield electrode 33. The inner shield electrode 33 is an electrically grounded stainless steel tube which is closed at the end that faces the  
30 ionization needle 15. This inner electrode 33 surrounds, and shields, an electrically isolated conductor 35 passing into its center. This innermost conductor 35 (i.e. the

ion collector electrode) is a collector for ions, and is connected to a fast current amplifier or electrometer 36 (e.g. Keithly model 428) and a digital storage oscilloscope 37 (e.g. LeCroy model 9450).

In the system shown in Figures 6A and 6B, the ions which surround the inner electrode 33 are forced inwards by a pulsed voltage. These ions travel from the FAIMS analyzer region 34 to the innermost conductor 35 through a series of 50  $\mu\text{m}$  holes 38 drilled through the inner shield electrode 33. The holes drilled in the inner shield electrode 33 are positioned about 2 cm from the end facing the ionization needle 15, and are spaced about 0.5 mm apart for a distance of 10 mm on one side of the inner shield electrode 33. The holes 38 drilled in the inner shield electrode 33 are located in this manner to minimize the variability in distance between the inner shield electrode 33 and the outer cylinder 32 in the vicinity of these holes 38. It was the inventors' objective to measure the ion abundance radial profiles of the ions located in the annular space (i.e. the FAIMS analyzer region 34) between the inner shield electrode 33 and the outer electrode 32 by pulsing the ions toward the inner shield electrode 33 and through the holes 38 and against the innermost ion collector electrode 35. The time-dependent distribution of ions arriving at the innermost conductor 35 is related to the physical radial distribution of ions around the inner electrode 33. Excessive variation in the distance between the two cylinders 32, 33 would have increased the uncertainty of the ion arrival times at the innermost conductor 35, thus decreasing the spatial resolution of the measurements made with this device.

Now referring to Figure 7, the high voltage, high frequency asymmetric waveform  $V(t)$ , applied to the FAIMS-R1-prototype of Figures 6A and 6B, is shown. The waveform is divided into two parts, the focussing period and the extraction period. The waveform was synthesized by an arbitrary waveform generator (e.g. Stanford Research Systems model DS340, not shown) and amplified by a pulse generator (e.g. Directed Energy Inc., model GRX-3.0K-H, not shown). The frequency of the waveform, and the relative duration of the high and low voltage portions of the waveform could easily be modified. Because of the high voltages, and steep rise-times of the square waves applied to this FAIMS-R1-prototype 30, the power consumption limits were severe, and waveforms in excess of about 1330 pulses (16

- 16 -

ms at 83,000 Hz) could not be delivered by this system without overheating electronic components of the high voltage pulse generator.

Note that, in the case of the FAIMS-R1-prototype 30, the high voltage, high frequency asymmetric waveform was applied to the outer cylinder 32 of the FAIMS-R1-prototype 30 shown in Figures 6A and 6B. Since all other forms of FAIMS discussed in this disclosure have the waveform applied to the inner tube or electrode, confusion may arise from the "polarity" of the waveform and the polarity of CV. In the FAIMS-R1-prototype 30 shown in Figures 6A and 6B, ions of type A (shown in Figure 1) are focussed during application of the opposite polarity waveform and CV than that shown for the devices in Figures 3A, 3B, 5A and 5B. Nevertheless, for simplification, the polarity will be written to be the same as if the device was constructed in the same way as those of the more conventional configuration. In other words the ions transmitted during application of waveform #1 will appear with DV positive and with CV negative. (Please note, however, that the actual voltages used on the device in Figures 6A and 6B are DV negative and CV positive).

As was observed in the conventional parallel plate FAIMS apparatus described earlier (Figure 2), the application of a high voltage asymmetric waveform  $V(t)$  will cause ions to migrate towards one of the FAIMS electrodes 2, 4 because of the changes in ion mobility at high electric fields (shown in Figures 1 and 2). This migration can be stopped by applying an electric field or compensation voltage CV in a direction to oppose the migration. For the FAIMS-R1-prototype 30 of Figures 6A and 6B, this CV was applied to the same electrode as the high voltage asymmetric waveform (i.e. the outer electrode 32), and was added to the waveform as a small dc bias (up to  $\pm 50$  V). At an appropriate combination of DV, and compensation voltage CV, a given ion will pass through the FAIMS device 30. The unit therefore acts like an ion filter. It is possible to fix conditions such that a single type of ion is isolated in the FAIMS analyzer 34 although a mixture flows uniformly out of the exit of the FAIMS device 30 although a mixture of ions are presented to the inlet of the FAIMS analyzer region 34.

The second part of the waveform shown in Figure 7 (i.e. the extraction period) was used to pulse the ions out of the FAIMS analyzer region 34 between the

60 Hz  
CV  
+ DV  
- 9.17 V  
to 32

- 17 -

outer electrode 32, and the inner shield electrode 33 (shown in Figures 6A and 6B). At the end of the focussing period, i.e. after 16 ms of waveform, the asymmetric waveform was replaced by a constant dc bias of approximately +30 V. This caused the ions from the annular space 34 between the outer electrode 32 and the inner shield electrode 33 to move in the direction of the inner shield electrode 33. A detector bias of -5 V, applied to innermost ion collector electrode 35, helped to carry the ions from the vicinity of the holes 38 in the inner shield electrode 33, through the holes 38 and into contact with the innermost ion collector electrode 35. The +30 V bias created an electric field of approximately 150 V/cm across the FAIMS analyzer region 34 and most ions located within this region 34 travelled across the 2 mm space in about 1 ms. The ion current due to the arrival of ions at the center inner shield electrode 33 can be predicted. For example, if only one type of ion, with mobility of  $2.3 \text{ cm}^2/\text{V-s}$ , e.g.,  $(\text{H}_2\text{O})_n\text{H}^+$  at ambient temperature and pressure conditions, was located in the FAIMS analyzer region 34, and if this ion was distributed evenly in the space, an approximately square-topped signal lasting approximately 0.6 ms should be observed. Deviation from this expected ion arrival profile would suggest that the ions were distributed in non-uniform profile across the FAIMS analyzer region 34 between the outer and inner cylinders of the FAIMS device 30.

Use  
20 Still referring to Figures 6A, 6B, and 7, the FAIMS-R1-prototype 30 was operated as follows. A 2L/min flow of purified air, Carrier Gas In (Cin), was passed into the cylinder 31 housing the ionization needle 15. Approximately 2000 V was applied to the needle 15, and the voltage was adjusted to produce a stable ionization current. The high voltage asymmetric waveform  $V(t)$  was applied to the outer FAIMS cylinder 32 for approximately 16 ms; this was followed by a 2 ms extraction pulse (Figure 7). The ion current striking the innermost ion collecting electrode 35 was detected and displayed on a digital oscilloscope 37. A measurement would typically consist of 100 averaged spectra, collected at a rate of approximately 5 Hz. Many experimental parameters were varied, including gas flow rates, the voltages of the asymmetric waveform  $V(t)$ , the dc voltage applied to the outer electrode CV, and the extraction voltage.

Figure 8 illustrates the ion arrival times at the innermost ion collector

electrode 35 observed by conducting these experiments. Each trace was recorded with 2500 V applied DV, but with variable CV voltages. As can be seen, during application of DV and CV, the radial distribution of ions is not uniform across the annular space of the FAIMS analyzer region 34. For example, at CV near -11 V, the ions are focussed into a narrow band near the inner electrode 33, and therefore are detected as a high intensity pulse occurring very early after the extraction voltage has been applied. At low CV, for example at -5.6 V, the ions are much more uniformly distributed between the walls of the concentric cylinders 32 33 making up the FAIMS analyzer region 34. When no electrical voltages are applied to the cylinders 32, 33, the radial distribution of ions should be approximately uniform across the FAIMS analyzer region 34 (data for this no-voltage experimental condition is not shown in this document). The experimental data shown in Figure 8 is evidence that the ion focussing is indeed occurring in FAIMS instruments. This focussing results in the ions being focussed in a uniform "sheet" or band around the inner cylinder 33 within the FAIMS analyzer region 34. As mentioned previously, to the inventors' knowledge, this focussing effect has never been observed or explained previously.

#### *Modes of Operation of FAIMS*

The focussing and trapping of ions by the use of asymmetric waveforms has been discussed above. For completeness, the behaviour of those ions which are not focussed within the FAIMS analyzer region will be described here. As explained earlier, the ions which do not have the high field ion mobility properties suitable for focussing under a given set of DV, CV and geometric conditions will drift toward one or another wall of the device, as shown in Figure 2. The rapidity with which they move to the wall depends on the degree to which their  $K_h/K$  ratio differs from that of the ion that might be focussed under the selected condition. At the very extreme, ions of completely the wrong property i.e. type A ion versus type C ion shown in Figure 1, will be lost to the walls very quickly.

The loss of ions should be considered one more way. If an ion of type A (Figure 1) is focussed at DV 2500 volts, CV -11 volts in a given geometry (for example, the FAIMS-E device of Figures 3A-3B), is it reasonable to expect that the ion will also be focussed if the polarity of DV and CV are reversed, i.e. DV of -2500

- 19 -

volts and CV of +11 volts (both applied to the inner electrode). It would seem that the reversal of polarity is a trivial exercise and the ion should be focussed, however, this is not observed. Instead, the reversal of polarity in this manner creates the mirror image effect of the ion focussing behaviour of FAIMS. The result of such polarity reversal is that the ions are not focussed, but rather are extremely rapidly rejected from the device, and collide with the walls of the cylinders 12, 13. The mirror image of a trapping valley, is a hill-shaped potential surface. The ions will slide to the center of the bottom of a trapping potential valley (2 or 3-dimensions), but will slide off of the top of a hill-shaped surface, and hit the wall of an electrode.

10 This apparently anomalous behaviour is a consequence of the cylindrical geometry of the FAIMS-E.

This is the reason for the existence, in the FAIMS, of the independent "modes" called 1 and 2. In this disclosure, the FAIMS instrument is operated in four modes: P1, P2, N1, and N2. The "P" and "N" describe the ion polarity, positive (P) and negative (N). The waveform (Figure 4, wave #1) with positive DV (where DV describes the peak voltage of the high voltage portion of the asymmetric waveform) yields spectra of type P1 and N2, whereas the reversed polarity (Figure 4, wave #2, negative DV) waveform yields P2 and N1. The discussion thus far has considered positive ions but, in general, the same principles can be applied to the negative ions, as explained in the preliminary note to the Detailed Description.

### *Separation Experiments*

Based on the FAIMS principles discussed above and the experiments conducted by the inventors to demonstrate the concept of ion focussing, the inventors have developed what is believed to be a previously unknown method for separation of isomers and different conformations of ions at substantially atmospheric pressure and substantially room temperature. Ion separation experiments involving several different types of ions are provided by way of example.

#### *A) Leucine/Isoleucine Separation*

In the prior art, there have been several attempts to distinguish the

amino acids leucine (Leu) and isoleucine (Ile) using sector, ion trap and quadrupole mass spectrometry. However, since Leu and Ile are structural isomers (i.e., identical elemental composition and molecular weight), their mass spectrometric differentiation has been limited to the interpretation of fragment ion mass spectra.

5 While differences in relative abundances of certain fragment ions can be used to unambiguously identify either of the two isomers in pure or pre-fractionated samples, mass spectrometric differentiation is incapable of selective determination within a mixture. In fact, the identification of the molecular ions of Leu and Ile, or those of their derivatives, has been demonstrated only following their

10 chromatographic separation prior to mass spectrometric analysis. The separation of these structural isomers has been achieved using ion exchange chromatography, high performance liquid chromatography, gas chromatography, and micellar electrokinetic chromatography. In general, these chromatographic methods require 5-15 minutes for separation and produce a 5-30 second transient pulse of analyte. The length of time

15 required for separation and the relatively short duration of the transient pulse of analyte limit this method of separation.

Based on the principles of high field asymmetric waveform ion mobility spectrometry, and based upon various experiments conducted by the inventors, a new and significantly improved method of separation of Leu and Ile has been

20 developed and is described here.

For this experiment, the FAIMS device coupled to a mass spectrometer as shown in Figure 5A was used. To generate negative ions, the electrospray needle was held at approximately -1900 V, giving an electrospray current of about 40 nA. The actual asymmetric waveform that was applied to the long inner cylinder of

25 FAIMS is shown in Figure 4 (Waveform #2). The maximum voltage of this waveform, referred to as the dispersion voltage (DV) was varied between 0 and -3300 V (which was the limit of the instrument). The frequency of the asymmetric waveform was constant at about 210 kHz. The CV, which was also applied to the long inner cylinder of the FAIMS analyzer, was scanned over specified voltage ranges.

30 As explained earlier, if the combination of DV and CV was appropriate, ions were not lost to the cylinder walls during their passage through the FAIMS analyzer and were transferred through an approximately 250  $\mu\text{m}$  orifice 19 to

the vacuum chamber of a mass spectrometer (PE SCIEX API 300 triple quadrupole). The MS orifice was electrically insulated from the FAIMS and a separate orifice voltage of -45 V was applied to it. Optionally, an offset voltage of -45 V was also applied to the entire FAIMS unit ( $V_{\text{FAIMS}}$ ) to enhance the sensitivity of the FAIMS-MS. The skimmer cone 18A of the MS was held at ground potential and the small ring electrode normally located behind the orifice of the API 300 was not incorporated into the present interface, resulting in some loss of sensitivity for low mass ions such as Leu and Ile. Compressed air was introduced into the carrier gas inlet ( $C_{\text{in}}$ ) at a flow rate of 3 L/min. Gas exited through the carrier gas outlet ( $C_{\text{out}}$ ) at 2 L/min and through the sample gas out port ( $S_{\text{out}}$ ) at 1 L/min. There was no flow through the "sample gas in" port ( $S_{\text{in}}$ ) in this study. The pressure inside the FAIMS analyzer was kept at approximately 770 torr.

*offset voltage*

For this experiment, commercially available samples of L-leucine and L-isoleucine were obtained. All standard solutions were prepared in 9:1 methanol/water (v/v) containing 0.2 mM ammonium hydroxide.

As explained earlier, FAIMS can be operated in any one of four modes, namely P1, P2, N1 or N2, where P and N describe ion polarity (positive and negative), and "1" and "2" are indicative of instrumental conditions. In general, low mass ions ( $m/z < 300$ ) such as Leu and Ile are transmitted in mode 1, while larger ions are transmitted in mode 2. In the present study, the ESI source was tuned to generate negative ions. Hence, all CV and mass spectra were collected using N1 mode. The asymmetric waveform used for N1 operation is shown in Figure 4 (Waveform #2).

The capability of FAIMS to separate ions generated from ESI of a mixture of Leu and Ile (5  $\mu\text{M}$  each) is shown in Figure 9. In each trace, (a) to (e), the voltage of the asymmetric waveform was set, and an ion-selected CV spectrum (IS-CV) was collected by scanning the CV while monitoring  $m/z \sim 130$ , the mass of the  $(M - H)^-$  ion of both compounds (where  $M = \text{C}_6\text{H}_{13}\text{NO}_2$ ). The dwell time and number of scans were kept constant for each spectrum. The IS-CV spectrum acquired without application of  $V(t)$  is shown in Figure 9 trace (a). Since the transmitted ions were not subjected to the electric fields caused by  $V(t)$  within the FAIMS analyzer they have experienced no change in mobility and appear in the

spectrum near  $CV = 0$  V. An increase to  $DV = -1700$  V, Figure 9 trace (b), results in the observation of three distinct peaks located at  $CV$  values of  $-0.1$  V,  $0.7$  V and  $1.3$  V. Increasing  $DV$  to  $-2300$  V, Figure 9 trace (c), caused most of the peaks in the IS-CV spectrum to shift to higher  $CV$  values, indicating significant changes in their high field mobility terms,  $K_h$ . The one exception is the first peak in spectrum at  $CV = -0.3$  V. This peak was identified as an amino acid dimer  $(M_2 - H)^+$  and is an N2 type ion, and as such is transmitted through FAIMS in N1 mode in a defocusing electric field. At  $DV = -2700$  V, Figure 9 trace (d), the single peak seen in Figure 4(c) at  $CV$  2.9 V has separated into two partially resolved peaks at  $CV$  values of 4.7 and 5.0 V. The separation of these two peaks may be improved by increasing  $DV$  to  $-3300$  V (the limit of the power supply), as shown in Figure 9 trace (e), giving  $CV$  values of 7.7 and 8.4 V. The position and separation of the group of smaller peaks at  $CV$  values of 2.9, 3.9 and 4.3 V in Figure 9 trace (e) also increased. The unambiguous assignment of any of these five peaks to Leu or Ile cannot be determined from this experiment. However, the mass spectrometer ensures that all of the peaks correspond to ions with  $m/z$  -130.

Identification of peaks in IS-CV spectra is analogous to using mass spectrometry as a selective detector for a chromatographic method such as CE, LC or GC. As with chromatography, it is necessary to have a set of matching standard solutions to identify the peaks in the CV spectrum. Determination of the identities of the peaks observed at  $CV$  values of 7.7 V and 8.4 V in Figure 9 trace (e) was accomplished by collecting IS-CV spectra ( $m/z$  -130,  $DV = -3300$  V) for 5  $\mu$ M solutions of either Leu or Ile. The spectra are shown in Figure 10 trace (a), where the dashed trace is the IS-CV spectrum of Leu and the solid trace is that of Ile. Note, Figure 10 trace (b) is the IS-CV spectrum of the mixture, as shown previously in Figure 9 trace (e), plotted over a narrower range of  $CV$  values. The peaks at  $CV$  values of 7.7 V and 8.4 V in the IS-CV spectrum shown in Figure 10 trace (a) may therefore be attributed to Leu and Ile, respectively.

The three small peaks seen in Figure 9 trace (e) at  $CV$  values of 2.9, 3.9 and 4.3 V were also present in the IS-CV spectra for both standard solutions. Identification of these peaks involved alternately tuning the FAIMS analyzer to a fixed  $CV$  value to selectively transmit one of the three ions, and collecting mass

spectra using varying FAIMS-MS sampling conditions. Gentle conditions, attained by reducing the collisional voltage in the MS interface, showed that these species correspond to the following adduct ions:  $(M(\text{CH}_3\text{O}(\text{CO}_2)))^-$  at CV 2.9 V;  $M(\text{CH}_3\text{COO})^-$  at CV 3.9 V; and  $M(\text{NO}_3)^-$  at CV 4.3 V where M is the neutral molecule of leucine or isoleucine. Under the more energetic conditions used to acquire the IS-CV spectra, these ion adducts were readily fragmented to yield the molecular ions  $(M - \text{H})^-$  of Leu and Ile at  $m/z$  -130.

As realized by the inventors, FAIMS, which continuously transmits one type of ion from a complex mixture is a significant improvement over conventional chromatographic methods of ion separation, especially when interfaced to relatively slow scanning mass spectrometers. As mentioned above, commonly used chromatographic methods for Leu and Ile are time-consuming (5-15 minute retention times) and result in narrow, finite impulses of analyte (5-30 seconds). The transient nature of these separation methods offers little flexibility in varying detection parameters and generally limits the degree to which the capabilities of the mass spectrometer may be exploited. With FAIMS, ion separation is independent of several experimental parameters associated with classical chromatography such as the stationary phase. Advantageously, problems encountered with the compatibility of LC and CE buffers (e.g. high salt content) and flow rates with the electrospray process are also eliminated.

In addition to its impressive separation ability, the FAIMS analyzer also functions to focus ions. (An experiment conducted by the inventors to show this focussing effect was described above.) A comparison of the observed ion current in Figure 9 trace (a) and 9 trace (e) shows an increase in signal of more than two orders of magnitude when DV is increased from 0 to -3300 V. This increase in ion current is attributed to the two-dimensional atmospheric pressure ion focusing discussed in detail earlier.

To illustrate the improvement in the mass spectra collected with FAIMS, spectra were acquired for the 5  $\mu\text{M}$  Leu/Ile mixture and are shown in Figures 11A-11C. The spectrum acquired with FAIMS disabled (DV = 0 V) and CV = 0 V is shown in Figure 11A. With no applied DV, there is no ion filtering effect, and hence no discrimination of the ions passing through the FAIMS analyzer. The mass

spectrum is complex, a commonly observed and often detrimental characteristic of electrospray mass spectra in the low-mass region. Peaks attributable to  $(\text{CO}_2(\text{CH}_3\text{O})^-; m/z -75)$ , oxalate ( $m/z -89$ ), Leu/Ile ( $\text{M-H}; m/z -130$ ),  $((\text{M}_2 - \text{H})^-; m/z -261)$  and  $((\text{Na}(\text{M} - \text{H})_2^-; m/z -283)$ , among others, are present. The peak  
5 observed at  $m/z -135$  is due to an impurity in the solvent or the ammonium hydroxide buffer. From this spectrum, the signal intensity for the dimer  $(\text{M}_2 - \text{H})^-$ , is roughly twice that of the molecular ions of Leu and Ile at a total analyte concentration of  $10 \mu\text{M}$ . At  $\text{DV} = -3300 \text{ V}$ , the mass spectra collected for the same sample mixture at CV values of  $7.7 \text{ V}$  and  $8.4 \text{ V}$ , i.e., the CV values of transmission  
10 of Leu and Ile, respectively, are simple and show one intense peak at  $m/z -130$  as shown in Figs. 11B and 11C. The FAIMS analyzer has effectively filtered out almost all of the background ions. This filtering action of the FAIMS analyzer was observed to improve signal to background ratios (S/B) for these analytes over spectra observed in conventional ESI-MS by at least a factor of 50.

15 The difference in the CV values required for the transmission of Leu and Ile is sufficient to permit selective monitoring of one of the species without interference from the other. This was illustrated by establishing response curves for both analytes present in a mixture. The response curve for Leu is shown in Figure 12. The total analyte concentration (i.e.,  $[\text{Leu}] + [\text{Ile}]$ ) in solution was kept constant at  
20  $2.500 \mu\text{M}$ , with the individual concentrations of each analyte varying over more than two orders of magnitude (i.e., from  $0.004$  to  $2.496 \mu\text{M}$ ). The plot is linear, ( $y=3.3[\text{Leu}]-0.01$ ,  $R^2= 0.9998$ ), indicating that there is no overlap of the Leu signal from Ile when it is monitored at  $\text{CV} = 7.7 \text{ V}$ , the value corresponding to its peak maximum. If an overlap of Leu and Ile was present at the CV values monitored, the  
25 Leu signal at low concentration would have been most notably affected, resulting in non-linearity in the response curve. An expanded view of an IS-CV plot acquired for the solution containing  $0.004 \mu\text{M}$  Leu and  $2.496 \mu\text{M}$  Ile, Figure 13, shows that the peaks are still well resolved at these concentrations. Note that the total concentration of  $2.500 \mu\text{M}$  was sufficiently low that no dimer ions were observed in  
30 the mass spectra. If the concentration of dimer ions had been significant, a negative deviation in the analytical curve at high analyte concentration would have resulted.

*B) Separation of Different Conformations of Gas-phase Molecular Ions using FAIMS*

Recently, considerable effort has been focused on attempting to understand the relationships between the amino acid sequence, the structure, and properties of proteins. A protein is composed of a series of linked amino acids, chemically covalently bonded to each other. Since there are about 20 different types of amino acids which can be included in this chain, the first level of the description of the structure of a protein is the listing of the names of these amino acids in the sequence that they appear in the protein. This is called the amino acid sequence.

Some of the amino acids have side groups which have the capability of creating chemical bonds to the side group of another amino acids someplace else in the amino acid sequence. This creates cross-linking. This cross linking is a very important structural element of proteins, because it forces certain areas of the protein sequence to be physically in close proximity to each other, in the final protein structure.

The chains of amino acids have the capability of forming small structures including loops, and hairpin shape structures that involve only a small number of amino acids. These structures are formed because some of the side chains of the amino acids interact weakly (non-covalently) with one another, and if the appropriate amino acids are in close proximity, then these weakly held structures will spontaneously form.

Finally, the combination of all of the smaller structures, and cross-links, give the protein an overall 3 dimensional structure. This structure is called the 'conformation'. This structure can be disrupted or modified many ways. The heating of the protein will 'denature' the protein. This usually means that the protein loses its functional capability because the 3-dimensional structure has been modified. This can occur because of the breaking of a cross-linking bridge, or the disruption of small or large scale structures via addition of thermal energy to the molecule.

Conformation, therefore, describes the 3-dimensional structure of the protein. The protein has a conformation whether or not the protein is capable of performing it's normal chemical activity, i.e. native, or denatured. Some terminology which describes the 3-dimensional structure may be 'extended', 'elongated', which

describes in a very non-specific way what we imagine the overall 3-dimensional structure will look like. Electrospray ionization (ESI), described above, has enabled the formation of intact gas-phase pseudo-molecular ions from large molecules, such as proteins. By coupling an ESI to a mass spectrometer (MS), ESI-MS has been used  
5 to provide information about conformations in solution. Since aqueous solutions at nearly physiological conditions are used in ESI-MS, this technique has been used to provide complementary structural information with other solution based methods such as Nuclear Magnetic Resonance (NMR).

In the prior art, information on gas-phase protein conformations has  
10 been gathered using several techniques. It has been proposed that these techniques can be divided into two general categories: chemical reactivity studies, and non-reactive studies. Chemical reactive studies, for example hydrogen-deuterium (H/D) exchange and proton transfer kinetics, examine the differences of reactivity of different conformations. Non-reactive studies use collisions with inert species to  
15 derive conformational information.

Based on the principles of high field asymmetric waveform ion mobility spectrometry discussed above, and based on experiments conducted by the inventors, a new and complementary method of measuring and studying different conformers of protein ions has been developed by the inventors and is described  
20 here.

Figure 5A shows a schematic view of a ESI-FAIMS-MS instrument of the type that was used in this study of different conformations. The electrospray needle  
15 and associated liquid delivery system were constructed by threading a 30 cm piece of fused silica capillary (50  $\mu\text{m}$  i.d., 180  $\mu\text{m}$  o.d.) through a 5 cm long stainless steel capillary (200  $\mu\text{m}$  i.d., 430  $\mu\text{m}$  o.d.), with the fused silica capillary protruding  
25 about 1 mm beyond the end of the stainless steel capillary. This stainless steel capillary, in turn, protruded about 5 mm beyond the end of a larger stainless steel capillary (500  $\mu\text{m}$  i.d., 1.6 mm o.d.) that was used for structural support and application of the high voltage. Solutions were delivered to the electrospray needle  
30 by a syringe pump (Harvard Apparatus model 22), at a flow rate of 1  $\mu\text{L}/\text{min}$ . For the generation of positive ions, the needle was held at approximately +2200 V giving an electrospray current of about 0.03  $\mu\text{A}$ . The electrospray needle was placed on the

center axis of the short inner cylinder, terminating about 5 mm short of the gap 11A between the two inner cylinders 11 and 12. The electrospray ions were driven radially outward by the electric field to the analyzer region through the 5 mm gap 11A between the two inner cylinders.

5           A high frequency (210 kHz), high voltage (0 to 4950 V p-p), asymmetric waveform (Figure 4) was applied to the long inner cylinder 12, thereby establishing the electric field between the inner and outer tubes. For the purposes of demonstrating separation of the conformers of the protein bovine ubiquitin all spectra were collected using P2 mode with DV = -3300 V. In addition to the high  
10 frequency waveform, a compensation voltage CV was also applied to the long inner cylinder 12. Although the CV can be scanned from -50 V to +50 V, the CV spectra herein are only shown from -12 V to 0V since the ions of bovine ubiquitin were transmitted through FAIMS within this CV window.

In this study, nitrogen gas was passed through a charcoal/molecular  
15 sieve gas purification cylinder and introduced into the FAIMS device through the carrier in ( $C_{in}$ ) port at a flow rate of 6 L/min. The gas exited through the sample out ( $S_{out}$ ) port at 1 L/min and through the carrier out ( $C_{out}$ ) port at 5 L/min. The sample in ( $S_{in}$ ) port was plugged in this study. A fraction of  $C_{in}$  was directed radially inward through the 5 mm gap 11A between the inner cylinders 11, 12, and  
20 acted to help desolvate the ions. While the ions formed by ESI were driven radially outward through the gap by the electric field, the inward flow of curtain gas prevented neutrals from entering the annular analyzer region. This portion of  $C_{in}$ , along with the neutrals, exited the FAIMS device *via* the  $S_{out}$  port.

The electrospray ions were carried by the gas stream along the length of  
25 the annular space between the outer cylinder and the long inner cylinder. If the combination of DV and CV was appropriate, and the ions were not lost to the tube walls, ions were transferred to the vacuum chamber of a mass spectrometer through the orifice 19 in the "sampler cone" 18 placed at the end of the FAIMS analyzer.

A custom interface was constructed for a tandem combination of  
30 FAIMS and a PE Sciex API 300 triple quadrupole mass spectrometer. The voltage of the sampler cone 18 was set to 44 V, whereas the skimmer cone 18A of the API 300 remained at ground potential for all experiments. The small ring electrode normally

located behind the orifice of the conventional API 300 interface was not incorporated into the new interface, resulting in some loss of sensitivity. Behind the skimmer cone, in a low pressure region ( $\sim 7 \times 10^{-3}$  torr), was an rf-only quadrupole (Q0) mass spectrometer 19A which acted to collisionally focus ions prior to their transmission  
5 into the first analyzer quadrupole (Q1). The voltage drop between the skimmer cone 18A and Q0 controlled the energy of the collisions in this region. The higher the voltage drop, the greater the extent of fragmentation. Unless otherwise stated, Q0 = -1 V was used.

Ion-selected CV spectra (IS-CV spectra) were obtained by scanning the  
10 compensation voltage applied to the FAIMS, while monitoring a single  $m/z$  value. "Total ion current" CV spectra (TIC-CV spectra) show the sum of the signal for all detected ions in a given  $m/z$  range as CV was scanned. The mass spectrum collected at fixed values of DV and CV revealed the identity of any ions transmitted through the FAIMS under those conditions.

15 The ubiquitin ions described and used in this experiment behave as type C ions, as shown in Figure 1. Analogous to the description of the behavior of positive type A ions, provided earlier, and referring back to Figures 1 and 2, a type C ion will travel further during the negative portion of the waveform ( $d_1 < d_2$ ), and will migrate toward the upper plate.

20 As explained earlier, there are a total of four modes of operation for FAIMS. The waveform with negative DV (Figure 4B) yields spectra of types P2 and N1, whereas the reversed polarity waveform yields P1 and N2 type spectra. In general, low mass ions ( $m/z$  is usually less than 300) are of type A (Figure 1) and are detected in mode 1, while larger ions, including the positively charged ubiquitin ions  
25 studied here, are type C ions and are detected in mode 2 (i.e. mode P2).

Now referring to Figure 14A, an ESI-FAIMS-MS mass spectrum is shown for a solution of 5  $\mu$ M bovine ubiquitin in 50/50/0.05 methanol/water/acetic acid (v/v/v) collected with FAIMS disabled. The solvent combination was selected for this illustration because several charge states are present. This mass spectrum  
30 essentially represents a conventional ESI-MS spectrum with somewhat lower sensitivity. Figures 14B-14E are ESI-FAIMS-MS spectra of the same solution used for Figure 14A with FAIMS in operation (DV = -3300V): Figure 14B shows a TIC-CV

spectrum ( $m/z$  30 to 2300), collected by scanning the CV from -12 V to 0 V, while Figures 14C-14E are mass spectra taken at specified CV values. The TIC-CV spectrum in Figure 14B shows two distinct peaks with maxima at CV = -8.0 V and CV = -5.4 V. Figures 14C-14E illustrate ESI-FAIMS-MS mass spectra of protein ions, taken at CV values indicated by the arrows in Figure 14B. The mass spectrum collected at CV = -8.0 volts (Figure 14C) is dominated by the  $[M + 7H]^{7+}$  ion of bovine ubiquitin. Unlike the conventional ESI-MS spectrum, Figure 14A, charge states higher than +7 are virtually absent in this mass spectrum. A mass spectrum (Figure 14D) taken at the CV corresponding to the small shoulder off of the main peak, at CV = -7.0 V (Figure 14B), shows a very different charge state distribution than Figure 14C. The  $[M + 6H]^{6+}$  ion is the most abundant ion in this mass spectrum and several higher charge states are also observed. Finally, the mass spectrum at CV = -5.4 volts (Figure 14E) shows yet another charge state distribution which is quite unlike the previous two mass spectra. In this spectrum, the  $[M + 7H]^{7+}$  ion is significantly reduced and the higher charge states (along with charge state +5) have increased in intensity. The increase in sensitivity in the mass spectra collected with the FAIMS operating (Figures 14C-14E), compared with the mass spectra with the FAIMS disabled (Figure 14A), is a result of the atmospheric pressure ion focusing mechanism discussed earlier.

The changes in the charge state distribution of the mass spectra collected at different CV values suggested that ion separation in FAIMS was sensitive to the structure of the protein ion. Consequently, IS-CV spectra for the individual charge states of bovine ubiquitin were collected using mass spectrometry, for the same solution as used to collect the data shown on Figure 14A, for charge states +5 to +13.

Now referring to Figure 15, some charge states (e.g., +10) show only one peak while others show multiple peaks (e.g., +8). Since the CV values for charge state +7 ( $m/z$  1224.6) are more negative than the CV values for charge state +6 ( $m/z$  1426.6) which are in turn more negative than the CV values for charge state +10 ( $m/z$  857.5), the position of the peak in an IS-CV spectrum is clearly not a function of  $m/z$ . The multiple peaks which appear in several of the IS-CV spectra in Figure 15 are attributed to coexisting, and distinct conformations of bovine

ubiquitin. In the following discussion we will illustrate the behavior of several of the conformers of bovine ubiquitin, as a function of solution phase conditions including the solvent composition and the pH.

Effect of Acid:

5 Different concentrations of acetic acid were used with 5  $\mu$ M solutions (55:45 H<sub>2</sub>O/MeOH v/v) of bovine ubiquitin to observe changes in the mass spectra and the CV spectra as a function of pH. The composition of the solvent has been selected to yield suitable mixtures of conformations for illustrating the effects of pH in the experiments discussed below. This solvent mixture was used for the data  
10 appearing in Figures 16 to 19. Figures 16A, 16C, and 16E, show mass spectra collected with the FAIMS disabled for 3 acetic acid (HOAc) concentrations from 0.04% to 4%, while Figures 16B, 16D, and 16F show the corresponding TIC-CV spectra ( $m/z$  from 30 to 2300) collected with the FAIMS in operation (DV = -3300V). At 0.04% acetic acid Figure 16A, the higher charge states are present at very  
15 low abundances (with 60% H<sub>2</sub>O by volume these charge states are no longer observed). This spectrum is very similar to one reported in an earlier study for bovine ubiquitin in its native state.

The TIC-CV spectrum in Figure 16B for the 0.04% acetic acid solution shows a strong peak at CV = -8.2 V with a small shoulder at CV = -6.0 V. The mass  
20 spectrum (FAIMS disabled) collected using a solution containing 0.4% HOAc (Figure 16C) shows a second charge state distribution, centered around  $[M + 12H]^{12+}$ , in addition to the distribution centered at  $[M + 7H]^{7+}$ . This second distribution is consistent with spectra collected for bovine ubiquitin in its denatured form as reported in an earlier study. The TIC-CV spectrum in Figure 16D for the 0.4% HOAc  
25 solution has a second peak at CV = -5.8 V, presumably due to the presence of the denatured bovine ubiquitin. Finally, at 4% acetic acid, the low sensitivity ESI-MS spectrum (Figure 16E) shows almost exclusively the higher charge states. The TIC-CV spectrum (Figure 16F) shows essentially one peak located at CV = -6.0 V. Consequently, it is clear that the changes in the TIC-CV spectra collected with the  
30 FAIMS in operation (DV = -3300 V) reflect changes seen in the mass spectra of bovine ubiquitin with the FAIMS disabled.

Figures 17A-17I show IS-CV spectra for some of the charge states (i.e., +7, +8, and +9) of bovine ubiquitin as a function of pH. Traces were obtained using the same solutions as traces in Figures 16A-16F. The CV spectra of some of the charge states, such as +10 to +13, did not show significant changes as a function of pH and were excluded from the Figure.

Figures 17A-17C are the IS-CV spectra of the +7, +8, and +9 charge states, respectively, collected using 0.04% acetic acid. These spectra reflect conditions in which the bovine ubiquitin is essentially in its native state. For charge state +8 in Figure 17B, two main peaks are observed at CV ~ -9 V and ~ -7 V. With 0.4% acetic acid, in the same solution, Figure 17E, the IS-CV spectrum for charge state +8 changes significantly. The peak that is observed CV ~ -9V is now virtually absent from the spectrum and a new peak at CV ~ -5 V is visible. This change suggests that as the bovine ubiquitin begins to unfold the conformer of bovine ubiquitin, reflected by the peak at CV ~ -9 V, is no longer favorable. At a concentration of 4 % acetic acid, Figure 17H, the peak at ~ -9 V is now completely absent from the spectrum. Charge state +7, Figures 17A, 17D, and 17G, also experiences significant changes with increasing concentrations of acetic acid. At low acetic acid levels, Figure 17 A, the IS-CV spectrum is dominated by a peak located at CV ~ -8 V. However, when 4 % acetic acid is used, Figure 17G, the IS-CV spectrum shows a shift that now favors the conformer which is present at CV ~ -7 V. Figures 17C, 17F, and 17I show that charge state +9 did not show significant changes as a function of pH.

To investigate behavior of the conformers at even lower pH values, hydrochloric acid, HCl, was added to solutions of 5  $\mu$ M bovine ubiquitin (solutions were still 55% H<sub>2</sub>O by volume). IS-CV spectra at pH 2.8, 2.1, and 1.8 are shown in Figures 18A-18I. The top traces (Figures 18A, 18B, and 18C) were obtained at the same pH as the last experiment in Figures 17G, 17H, and 17I (i.e., pH ~ 2.8). This was done to permit a comparison based on changing the acids only. The IS-CV traces for charge states +8 and +9 at pH 2.8 (Figures 18B, and 18C) in HCl are very similar to that observed in Figures 17H, and 17I using acetic acid at pH 2.8. The IS-CV spectra for the other charge states not shown also gave similar results. However, the IS-CV spectrum for charge state +7 in HCl at pH 2.8 (Figure 18A)

shows significant differences from the IS-CV spectrum (Figure 17G) for charge state +7 at pH 2.8 in acetic acid. The former spectrum more closely resembles IS-CV spectra for acetic acid at slightly higher pH values, but with an additional peak at CV ~ -5 V that was not observed when using acetic acid.

5 By decreasing the pH from 2.8 (Figures 18A, 18B, and 18C) to 2.1 in Figures 18D, 18E, and 18F, there are significant changes that are observed for charge states +7, +8, and +9 respectively (and others). In all instances, a peak at a CV of between -5 and -6 V becomes favored. This trend is continued as the pH is lowered even further to 1.8 (Figures 18G, 18H, and 18I). The IS-CV spectra that were  
10 obtained for this solution were markedly more noisy and less intense. However, the dominance of the peak at the least negative CV value is still apparent.

The data illustrated in Figures 15 and 19 permit a comparison of the CV spectra of bovine ubiquitin at several charge states, using solutions of acetic acid at pH 3.4 (Figure 15, 50:50 water/MeOH) with CV spectra collected using a solution  
15 acidified to pH 2.1 with HCl (Figure 19, 55:45 water/MeOH). The traces for charge states +10 through +13, in Figure 19, are very similar to those observed in Figure 15. Changes in the CV spectra of charge states +5 through +9 were observed.

#### Effect of Solvent:

The CV spectra and mass spectra shown in Figure 20 illustrate the  
20 effect of changing the solvent mixture from 50:50 water/MeOH Figures 20A, 20C, and 20E, to 55:45 water/MeOH v/v Figures 20B, 20D, and 20F, while maintaining low acid (0.04% HOAc) concentration. The mass spectra (Figures 20A and 20B) were collected with the FAIMS disabled while the CV spectra (Figures 20C through 20F) were collected with the FAIMS in operation (DV = -3300V). As is consistent  
25 with previous studies, increasing amounts of organic solvent cause bovine ubiquitin to denature. This is shown by the slight increase in the higher charge states of mass spectra acquired with the solution containing the higher percentage of organic solvent (Figure 20A, 50% MeOH) relative to that shown for lower percent organic solvent (Figure 20B). The FAIMS CV spectra also reflect this change in solvent composition  
30 as indicated in the TIC-CV spectra in Figures 20C and 20D. Furthermore, the IS-CV spectrum for charge state +8 (Figures 20E, and 20F) also undergoes a significant

change over this solvent range. The peak at -9 V in the CV spectrum of the +8 charge state collected using a solution containing 50% MeOH (Figure 20E) is much less intense than that observed at -9 V for a solution which contained 45% MeOH (Figure 20F).

5    Exchange Reactions:

Mass spectra of the three distinct peaks observed in the CV spectrum of charge state +8 (Figure 21A) were collected to investigate differences among the conformers of this charge state. The CV spectrum (Figure 21A) and the mass spectrum collected for each peak, for a solution containing 5  $\mu$ M bovine ubiquitin in 10 50% H<sub>2</sub>O and 0.04% acetic acid are shown in Figures 21C, 21E, and 21G. The expanded views of the mass spectra show that several sodium ions have replaced protons,  $(M+nH+mNa)^{+n+m}$  where  $n+m=8$ , in each conformation of the +8 charge state of bovine ubiquitin. The mass spectrum of the peak at CV = -4.8 V (Figure 21G) clearly shows up to four sodium replacements, the mass spectrum of the peak 15 at CV = -8.9 V (Figure 21C) shows virtually no proton replacement, and the mass spectrum of the peak at CV = -6.9 volts (Figure 21E) shows that up to two sodium ions have replaced protons in the +8 ion. We speculate that the differences in the aqueous phase, three-dimensional structures of the conformers of the +8 charge state of the bovine ubiquitin ion have resulted in varying degrees of replacement of protons 20 by sodium ions in the gas-phase conformer.

It could be argued that the location of these adducts in the CV spectrum is not a *consequence* of the different conformations of bovine ubiquitin, but instead that the sodium adduct actually *causes* the ions to be located in the CV spectrum as shown in Figure 21A. Previous results with ESI-FAIMS-MS obtained by the inventors 25 have shown that for a smaller species (e.g., leucine enkephalin, MW 555.5), adduct ions can alter the CV of a given species in the compensation voltage spectrum. Consequently, two additional experimental results are described to show that the presence of different degrees of sodium replacement is a result of the different conformations, and that the location of the ion in these CV spectra is not altered by 30 replacement of protons with sodium ions.

The spectra in Figures 21B, 21D, 21F, and 21H were obtained after 1

mM of sodium chloride was added to the solution that was used to collect the data for Figures 21A, 21C, 21E, and 21G. Note that due to the distribution of the sodium-containing ions over several  $m/z$  values, the IS-CV spectrum collected in Figure 21B was collected over the  $m/z$  range 1071-1080 rather than at  $m/z$  1071.6 exclusively. The IS-CV spectra in Figures 21A and 21B are very similar, and the CV values of the three peaks corresponding to +8 of bovine ubiquitin have not changed significantly. However, the expanded views of the mass spectra collected for these three species have changed. Furthermore, the mass spectrum for the species at CV = -4.8 V with no sodium added (Figure 21G) is very similar to the mass spectra for the species at CV = -6.9 V with 1 mM sodium added (Figure 21F). The question to be asked is whether the number of sodium ions bound to the protein ion determine its position in the CV spectrum or does the conformation of the protein ion determine the degree of sodium substitution. If the replacement of protons with sodium ions was causing the CV shift, these two mass spectra should be observed at virtually identical locations in the CV spectra. Thus, the number of sodium replacements reflects structural differences in the conformers represented by the three peaks observed for the +8 state of bovine ubiquitin.

An IS-CV spectrum for the +6 charge state (Figure 22A) is dominated by a peak at CV = -7.5 V, however there is also a smaller peak at CV = -5.7 V. Figures 22B and 22C show mass spectra that were collected for these peaks at CV = -5.7 V and CV = -7.5 V (note Figure 22B represents a sum of 50 mass spectra). For this charge state, the ion which is amenable to higher level of sodium replacement of protons is observed at the more negative CV of the two conformers. Conversely, the data for the +8 charge state of bovine ubiquitin (Figure 21), indicates that the conformer which accumulates the higher number of sodium ions was observed at less negative CV values.

The effect of other species, beside sodium, on CV spectra for bovine ubiquitin was also investigated. Figure 23 shows ESI-FAIMS-MS data collected using an excess of potassium di-hydrogen phosphate added to a solution of 5  $\mu$ M bovine ubiquitin in 50%  $H_2O$  and 0.02% acetic acid. Figure 23A shows an IS-CV spectrum for the +8 charge state of bovine ubiquitin that is very similar to that observed previously in this study for solutions of 50%  $H_2O$ , Figure 21A. The change in the

relative abundances of the conformers at CV ~ -9 V and ~ -5 V compared with, for example, Figure 21A, can be attributed to the change in pH caused by the difference in HOAc concentration, and added  $\text{KH}_2\text{PO}_4$ . Expanded views of the mass spectra, obtained at CV values -8.9 V, -6.9 V, and -4.8 V in Figure 23A, are shown in Figures 23B, 23C, and 23D, respectively, and show significant differences in the abundances of phosphate adducts. As was observed with sodium, the degree of adduct formation is a function of the conformer. However, the results in Figure 23 show the opposite trend compared with the results shown in Figure 21. That is, in Figure 23, the conformer that appears in the CV spectrum at CV ~ -9 V (Figure 23B) shows the most intense phosphate adduct ion whereas the conformer that appears in the CV spectrum at CV ~ -5 V shows the least intense phosphate ion adduct (Figure 23D). Note also that the addition of phosphate did not modify the CV at which each of the three conformers appeared (compare to Figure 21A). We also note that none of the three conformers of the +8 ion had a tendency to replace protons with potassium ions.

These examples illustrate that by examining 'spectator' ions such as sodium and phosphate, we can gain clues into the differences in the individual conformations. Experiments have indicated that the replacement/addition of the 'spectator' has not affected the ion conformation. It is possible that the formation of some replacements/additions may lead to significant changes in the conformation of an ion. Such a change would be expected (in most cases) to be reflected by peak shifts in the IS-CV spectra.

#### Concentration Effects:

In an earlier study, the inventors observed more than one peak in an IS-CV spectrum of a solution of leucine enkephalin ( $m/z$  556.5) with ESI-FAIMS-MS (see R. Purves and R. Guevremont, *Anal. Chem.* 71, 2346 (1999), R. Guevremont and R. Purves *J. Am. Soc. Mass Spectrom.* 10, 492 (1999)), and attributed the additional peaks to a series of cluster ions of the type  $(nM+nH)^{n+}$ . To ensure that the different peaks observed in the IS-CV spectra in this study (e.g. Figure 15) were not a consequence of the formation of multimers or other cluster ions, different concentrations of bovine ubiquitin were studied. For the concentration range from 1

μM to 100 μM bovine ubiquitin (in 50% water and 0.04% acetic acid), no significant changes in the shapes of the IS-CV spectra for any charge state were observed. If the multiple peaks in an IS-CV spectrum were caused by the formation of cluster ions, the relative amounts of the various peaks in an IS-CV spectrum would change as a function of concentration.

Comparison with Drift Tube Mobility Spectrometry

ESI-drift tube mobility spectrometry/MS has been used to examine the conformations of a number of proteins. The separation of conformers in a drift tube is based on the ion cross section, whereas the separation of ions in FAIMS is based on heretofore unknown properties of the ions. It is expected therefore, that there will be many similarities, and also significant differences in the array of conformations detected by these two, independent approaches.

Valentine *et. al.* have used their ESI-drift tube mobility spectrometry/MS, and proton transfer reagents, to study the conformers of bovine ubiquitin (S.J. Valentine, A.E. Counterman, D.E. Clemmer, J. Am. Soc. Mass Spectrom. 8, 954 (1997)). In their work, the authors divided their results for collision cross section into three types of conformations: elongated, partially folded, and compact. Figure 24 summarizes the conformational information obtained for bovine ubiquitin by ESI-FAIMS-MS as a plot of CV values for each of the peaks appearing in the CV spectra for each of the individual charge states. The "low acid" solution contained 55% H<sub>2</sub>O and 0.04% acetic acid while the "high acid" solution contained 55% H<sub>2</sub>O and the pH adjusted to 2.1 with HCl. The peak maxima that were observed in the IS-CV spectra were classified as being observed either in the "low acid" solution, the "high acid" solution, or both.

For some charge states in Figure 24 (i.e., +10 through +13), the CV remains approximately constant between -5 and -6 V independent of the acid concentration (pH). For the remaining charge states (i.e., +5 through +9), there are several resolved conformers, and the relative abundances of the conformers is dependent on the solution conditions. The most negative CV values are observed for conformers only present in the "low acid" solution. For charge states +5 through +8, Valentine *et. al.* reported the co-existence of elongated and partially folded

- 37 -

conformations. Consequently, both FAIMS, and drift tube mobility spectrometry techniques identified the same charge states having a multiplicity of conformations.

Although the invention has been described and illustrated with reference to specific embodiments, those skilled in the art will recognize that the invention may  
5 be otherwise embodied within the scope of the following claims. Specifically, while the experiments describe the use of a mass spectrometer to measure transmitted ions at an ion outlet of a FAIMS device, it will be understood that a mass spectrometer is not necessary once the values for DV, CV and other operating conditions for the FAIMS device have been determined for separating a desired ion. Thus, collection of  
10 desired transmitted ions may occur at the ion outlet of a FAIMS device, rather than entering a mass spectrometer for measurement.

## WE CLAIM:

1. A method for identifying ions having substantially the same mass to charge ratio but having different ion mobility characteristics, comprising the steps of:
  - a) providing at least one ionization source of ions;
  - 5 b) providing an analyzer region defined by a space between at least first and second spaced apart electrodes, said analyzer region being in communication with a gas inlet, a gas outlet, an ion inlet and an ion outlet, and introducing said ions into said analyzer region through said ion inlet;
  - 10 c) applying an asymmetric waveform voltage and a direct current compensation voltage to at least one of said electrodes;
  - d) setting said asymmetric waveform voltage;
  - e) varying said direct current compensation voltage and measuring resulting transmitted ions at said ion outlet, so as to produce a compensation voltage scan for said transmitted ions; and
  - 15 f) identifying peaks in said compensation voltage scan.
2. The method claimed in claim 1, further comprising the step of setting said direct current compensation voltage to correspond to one of said peaks to  
20 separate a desired ion from other ions with substantially the same mass to charge ratio.
3. The method claimed in claims 1 or 2, which includes operating substantially at atmospheric pressure and substantially at room temperature.
4. The method claimed in claim 1, 2 or 3, which includes generating said ions for said source of ions by electrospray ionization.
- 25 5. The method claimed in any preceding claim, which includes detecting said transmitted ions by mass spectrometry.

6. The method claimed in claim 5, which includes subjecting the transmitted ions to a mass analysis scan to provide ion intensity data over a selected range of mass to charge ratios.
7. The method claimed in any preceding claim, which includes providing  
5 a gas flow through said analyzer region, so as to transport said ions along said analyzer region.
8. The method claimed in claim 1, which includes generating ions from isomers as said source of ions.
9. The method claimed in claim 8, which further includes selecting said  
10 isomers from the group comprising stereoisomers, enantiomers, diastereomers, constitutional isomers, geometrical isomers, positional isomers, and cis-trans isomers.
10. The method claimed in claim 8, wherein, said isomers are leucine and isoleucine.
11. The method claimed in claim 1, which includes generating ions from  
15 conformers as said source of ions.
12. The method claimed in claim 11, which further includes selecting said conformers from one of synthetic polymers and biological polymers.
13. The method claimed in claim 12, wherein, said conformers are  
20 different conformers of bovine ubiquitin.
14. The method as claimed in claim 2, which includes collecting the desired ions for further processing.
15. A method for identifying constitutional isomers, comprising the steps

of:

- a) providing at least one ionization source of at least one constitutional isomer;
- b) providing an analyzer region defined by a space between at least first and second spaced apart electrodes, said analyzer region being in communication with a gas inlet, a gas outlet, an ion inlet and an ion outlet, and introducing said ions into said analyzer region through said ion inlet;
- c) applying an asymmetric waveform voltage and a direct current compensation voltage to at least one of said electrodes;
- d) setting said asymmetric waveform voltage;
- e) varying said direct current compensation voltage and measuring resulting transmitted ions at said ion outlet, so as to produce a compensation voltage scan for said transmitted ions; and
- f) identifying peaks in said compensation voltage scan corresponding to a constitutional isomer.

16. The method claimed in claim 15, further comprising the step of setting said direct current compensation voltage to correspond to a peak so as to separate a desired constitutional isomer.

17. The method claimed in claim 15 or 16, which includes operating substantially at atmospheric pressure and substantially at room temperature.

18. The method claimed in claim 15, 16 or 17, which includes detecting said transmitted ions by mass spectrometry.

19. The method as claimed in claim 16, which includes collecting the desired constitutional isomers for further processing.

20. The method claimed in any one of claims 15-19, which includes providing a gas flow through said analyzer region, so as to transport said ions along

said analyzer region.

21. A method for separating one of leucine ions and isoleucine ions from a gaseous mixture of leucine and isoleucine ions, comprising the steps of:

- a) providing a solution of leucine and isoleucine;
- 5 b) producing a gaseous mixture of leucine and isoleucine ions from said solution of leucine and isoleucine ions;
- c) providing an analyzer region defined by a space between at least first and second spaced apart electrodes, said analyzer region being in communication with a gas inlet, a gas outlet, an ion inlet and an ion outlet, and introducing said leucine and isoleucine ions into said  
10 analyzer region through said ion inlet;
- d) applying an asymmetric waveform voltage and a direct current compensation voltage to at least one of said electrodes;
- e) setting said asymmetric waveform voltage; and
- 15 f) setting said direct current compensation voltage to separate one of said leucine and isoleucine ions from said ion mixture.

22. The method claimed in claim 21, which includes operating substantially at atmospheric pressure and substantially at room temperature.

23. The method claimed in claim 21 or 22, which includes providing a gas  
20 flow through said analyzer region, so as to transport said ions along said analyzer region.

24. The method claimed in claim 21, which includes generating negatively charged leucine and isoleucine ions.

25. A method for identifying different conformers, comprising the steps  
25 of:


- a) providing at least one ionization source for producing a sample of

- different conformers;
- 5      b) providing an analyzer region defined by a space between at least first and second spaced apart electrodes, said analyzer region being in communication with a gas inlet, a gas outlet, an ion inlet and an ion outlet, and introducing said sample into said analyzer region through said ion inlet;
- 10      c) applying an asymmetric waveform voltage and a direct current compensation voltage to at least one of said electrodes;
- 10      d) setting said asymmetric waveform voltage;
- 10      e) varying said direct current compensation voltage and measuring resulting transmitted ions at said ion outlet, so as to produce a compensation voltage scan of said transmitted ions; and
- 10      f) identifying peaks in said compensation voltage scan corresponding to different conformers.
- 15      26. The method claimed in claim 25, further comprising the step of setting said direct current compensation voltage to substantially correspond to a peak so as to separate a desired conformer.
- 20      27. The method claimed in claim 25 or 26, wherein, said analyzer region is operated substantially at atmospheric pressure and substantially at room temperature.
28. The method claimed in claim 25, 26 or 27, wherein, said sample is produced by electrospray ionization.
29. The method claimed in any one of claims 25-28, which includes detecting said transmitted ions by mass spectrometry.
- 25      30. The method claimed in claim 29, which includes subjecting the transmitted ions to a mass analysis scan to provide ion intensity data over a selected range of mass to charge ratios, so as to identify different conformations of


ions at given charge states.

31. The method claimed in any one of claims 25-30, which includes providing a gas flow through said analyzer region, so as to transport said ions along said analyzer region.


5 32. The method claimed in claim 25, wherein, said ions are conformers of bovine ubiquitin.



**FIG. 1**



**FIG. 2**



**FIG. 3A**





FIG. 3B



**FIG. 4**




FIG. 5A





FIG. 5B

8/26








FIG. 6B


FIG. 7


11/26

**FIG. 8**




SUBSTITUTE SHEET (RULE 26)

12/26



**FIG. 9**






**FIG. 10**

14/26


FIG. 11AFIG. 11BFIG. 11C

15/26




FIG. 12FIG. 13



17/26

**FIG. 15**







18/26

**FIG. 16A****FIG. 16B****FIG. 16C****FIG. 16D****FIG. 16E****FIG. 16F**

charge state = +7

charge state = +8

charge state = +9




(I) 0.04% acid  
pH = 4.0FIG. 17AFIG. 17BFIG. 17C(II) 0.4% acid  
pH = 3.4FIG. 17DFIG. 17EFIG. 17F(III) 4% acid  
pH = 2.8FIG. 17GFIG. 17HFIG. 17I

charge state = +7



charge state = +8

charge state = +9




pH = 2.8


FIG. 18AFIG. 18BFIG. 18C

pH = 2.1

FIG. 18DFIG. 18EFIG. 18F







pH = 1.8

FIG. 18GFIG. 18HFIG. 18I



**FIG. 19**

22/26

FIG. 20AFIG. 20BFIG. 20CFIG. 20DFIG. 20EFIG. 20F

23/26

no NaCl added




FIG. 21A

1 mM NaCl added




FIG. 21B

CV = -8.9 V



FIG. 21C




FIG. 21D

CV = -6.9 V




FIG. 21E



FIG. 21F

CV = -4.8 V




FIG. 21G








FIG. 21H


24/26

**FIG. 22A****FIG. 22B****FIG. 22C**


25/26




**FIG. 23A**




**FIG. 23B**



**FIG. 23C**



**FIG. 23D**

**FIG. 24**

# INTERNATIONAL SEARCH REPORT

<b>A. CLASSIFICATION OF SUBJECT MATTER</b> IPC 7 G01N27/64 H01J49/04 H01J49/42 B01D59/48		International Application No PCT/CA 99/00714
According to International Patent Classification (IPC) or to both national classification and IPC		
<b>B. FIELDS SEARCHED</b> Minimum documentation searched (classification system followed by classification symbols) IPC 7 G01N H01J B01D		
Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched		
Electronic data base consulted during the international search (name of data base and, where practical, search terms used)		
<b>C. DOCUMENTS CONSIDERED TO BE RELEVANT</b>		
Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
P, X	GUEVREMONT R ET AL: "High field asymmetric waveform ion mobility spectrometry-mass spectrometry: an investigation of leucine enkephalin ions produced by electrospray ionization" JOURNAL OF THE AMERICAN SOCIETY FOR MASS SPECTROMETRY, US, ELSEVIER SCIENCE INC., NEW YORK, NY, vol. 10, no. 6, page 492-501 XP004173039 ISSN: 1044-0305 the whole document. --- --- -/---	1, 15, 21, 25
<input checked="" type="checkbox"/> Further documents are listed in the continuation of box C.		
<input checked="" type="checkbox"/> Patent family members are listed in annex.		
* Special categories of cited documents : "A" document defining the general state of the art which is not considered to be of particular relevance "E" earlier document but published on or after the international filing date "L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified) "O" document referring to an oral disclosure, use, exhibition or other means "P" document published prior to the international filing date but later than the priority date claimed "T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention "X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone "Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art. "&" document member of the same patent family		
Date of the actual completion of the international search 22 December 1999		Date of mailing of the international search report 12/01/2000
Name and mailing address of the ISA European Patent Office, P.B. 5818 Patentlaan 2 NL - 2280 HV Rijswijk Tel. (+31-70) 340-2040, Tx. 31 651 epo nl. Fax: (+31-70) 340-3016		Authorized officer Hulne, S

## INTERNATIONAL SEARCH REPORT

International Application No

PCT/CA 99/00714

## C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT

Category	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
Y	<p>HUDGINS R R ET AL: "High resolution ion mobility measurements for gas phase proteins: correlation between solution phase and gas phase conformations" INTERNATIONAL JOURNAL OF MASS SPECTROMETRY AND ION PROCESSES, NL, ELSEVIER SCIENTIFIC PUBLISHING CO. AMSTERDAM, vol. 165-166, page 497-507 XP004103206 ISSN: 0168-1176 page 498 -page 499</p>	1,3,15, 18,21, 22,25, 27-29
Y	<p>RIEGNER ET AL: "Qualitative evaluation of field ion spectrometry for chemical warfare agent detection" PROCEEDINGS OF THE 45TH ASMS CONFERENCE ON MASS SPECTROMETRY AND ALLIED TOPICS, June 1997 (1997-06), pages 473a-473b, XP000865529 cited in the application the whole document</p>	1,3,15, 18,21, 22,25, 27-29
A	<p>BURYAKOV ET AL.: "A new method of separation of multi-atomic ions by mobility at atmospheric pressure using a high-frequency amplitude asymmetric strong electric field" INTERNATIONAL JOURNAL OF MASS SPECTROMETRY AND ION PROCESSES., vol. 128, 1993, pages 143-148, XP000865595 ELSEVIER SCIENTIFIC PUBLISHING CO. AMSTERDAM., NL ISSN: 0168-1176 cited in the application figure 2</p>	1,15,21, 25
A	<p>CARNAHAN B. ET AL.: "Field ion spectrometry - a new analytical technology for trace gas analysis" PROCEEDINGS OF THE 41ST ISA ANALYSIS DIVISION SYMPOSIUM, vol. 29, 21 - 24 April 1996, pages 85-94, XP000863733 cited in the application figure 2</p>	1,15,21, 25
A	<p>US 5 420 424 A (CARNAHAN BYRON L ET AL) 30 May 1995 (1995-05-30) cited in the application abstract</p>	1,15,21, 25

# INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No  
PCT/CA 99/00714

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 5420424 A	30-05-1995	CA 2148166 A	30-10-1995
		EP 0679886 A	02-11-1995
		FI 951910 A	30-10-1995
		IL 113468 A	20-11-1997
		JP 8054373 A	27-02-1996